

Nurse Discharge Checklist



- Identified staff member will send Masimo invite from website to patient's email or electronic device.
Make sure to use CHROME
- Discharge Nurse ensures consent is completed and signed by patient/caregiver and witnessed
- Discharge Nurse assists with app download and set up on patient's mobile device
- Ensure patient's mobile device works
- Discharge Nurse enters patient information into Masimo system
- Ensure a total of **3 sensors** are provided to the patient. 1 placed on finger, 2 inside the box/kit
- Discharge Nurse provides patient education on remote pulse ox
- Masimo Pulse Ox is applied to patient on ring finger
- Instruct patient to change sensor when battery light turns **RED** or every 4 days
- Discharge Nurse reviews with patient discharge instructions sheet, Patient daily checklist, Light Indicator Guide and contact numbers
- If patient discharged home on Oxygen (MAX 2 Liters) contact Social Work and Case Management for Home Health Care arrangements
- Prior to discharge, call 786-533-4444 to confirm patient is set up and the device data is captured by Remote Monitoring Staff